



Certificate of Compliance

Certificate: 80069776

Master Contract: 301811

Project: 80069776

Date Issued: 2021-04-20

Issued To: Vitacore Industries
151-8518 Glenlyon Parkway
Burnaby, British Columbia, V5J 0B6
Canada

Attention: Yang Fei

The products listed below are eligible to bear the CSA Mark shown

Issued by: Wendy Zhu
Wendy Zhu, Ph. D., P. Eng.



PRODUCTS

CLASS - C720402 - OCCUPATIONAL HEALTH & SAFETY Surgical Non-Powered Air-Purifying Respirators

Surgical Respirator Model CAN99, single-use and disposable, 99PFE-L3, one size

APPLICABLE REQUIREMENTS

- The Code of Federal Regulation 42 CFR Part 84, Approval of Respiratory Protective Devices, Subpart K (Nonpowered Particulate Respirators).
- The Code of Federal Regulations, Title 16- Commercial Practices, Part 1610, Standard for the Flammability of Clothing Textiles
- ASTM F1862-17 Standard Test Method for Resistance of Medical Face Masks to Penetration by Synthetic Blood
- CSA TechNote TN202, Filtering Facepiece Respirator Head Strap / Harness Assembly Strength Requirements and Test Procedure



Supplement to Certificate of Compliance

Certificate: 80069776

Master Contract: 301811

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
80069776	2021-04-20	Original certification of Surgical Respirator Model CAN99, single-use and disposable, 99PFE-L3, one size. The Code of Federal Regulation 42 CFR Part 84, Subpart K The Code of Federal Regulations, Title 16- Commercial Practices ASTM F1862-17 CSA TechNote TN202